## AED /Bleed Kit Inspection Form

Inspected By:		Location of Inspection (Building):			
Date of Inspection			AED Serial Number:		
Use the following maintenance checklist Check the following Pass or Fail	when y	you	check y	our unit.	
Check the following	PAS	S	FAIL	Comments	
Is the green check showing that the unit is ready to use?					
Is the unit clean, undamaged, and free of excessive wear?					
Are there any cracks or loose parts in the housing?					
Verify that electrodes are within their expiration date.					
Verify electrodes are connected to the unit and sealed in their package. Replace if expired.				Exp. Date:	
Are all cables free of cracks, cuts, and exposed or broken wires?					
Turn the unit on and off and verify the green check indicates ready for use.					
Batteries within expiration date. Replace if expired.				Exp. Date:	
Check for adequate supplies (mask, gloves, extra batteries)					
Alarm on AED Box operates correctly					
If Bleeding Control Kit is present at location – Seal is not broken or missing		·			